EDTA WASH - CHELARA L - ENDOPREP L

(EDTA Chelating and Irrigation Liquid)

INSTRUCTIONS FOR USE (Please read carefully before use)

Doc. No: **HH/IFU/01** Approved By: **Mr. Vikram Vora** Language: **English**

Revision	01/01
Date	29.08.2026

PRODUCT DESCRIPTION: The main component, EDTA, forms a complex with dentine calcium after injection into dentine, loosening the structure of hard tooth tissues and easing enlargement of calcified and sclerous canal orifices, cleaning and forming hard-accessible tooth canals.

MATERIAL OF CONSTRUCTION: 17% Disodium EDTA, 0.7% Cetrimide, excipient q.s. ad 100 ml.

INDICATIONS: Irrigation and chelating agent in root canal treatment; detection of root canal orifice; chemical enlargement of root canals when preparing for filling.

CONTACT TIME: 2–3 minutes per application in the root canal.

PROCEDURE & PRODUCT HANDLING:

- The liquid is applied with the help of a syringe from the luer lock of the bottle, with a side-vent root needle or cotton swab. Put it into the pulp cavity to start mechanical enlargement.
- Then, using a root needle, apply the liquid into the root canal for 2–3 minutes.
- Repeat the procedure alternating chemical and mechanical enlargement with impregnation of dentine with a fresh portion of the liquid if necessary.

CAUTION / WARNING:

1. Device to be used by a qualified dentist or dental technician only.
2. Use with great care in allergic patients.

PRECAUTIONS:

1. Close the bottle cap tightly after use.

CONTRAINDICATIONS:

1. Drug idiosyncrasy.

INTERACTIONS WITH OTHER DENTAL MATERIALS: Kindly communicate and inform the manufacturer with complete details in case of any incidences of interaction not mentioned here.

ADVERSE REACTIONS: No known side effects when the product is applied correctly. Kindly communicate and inform the manufacturer in details of any instance.

PACKAGING: Liquid, 100 ml.

STORAGE CONDITIONS: Store at a temperature between 5°C to 25°C in a well-sealed container. Keep away from direct sunlight. Keep locked up and out of the reach of children. Close bottle cap tightly after use. Do not use after expiry date.

SHELF LIFE: 2 years from the date of manufacturing.

DISPOSAL: Dispose in accordance with local, state and federal environmental regulations.

This product is specially formulated for use in dentistry and to be used by qualified dental personnel.

LOT Symbol for “BATCH CODE”	Symbol for “USE BY”	Symbol for “NON STERILE”
Symbol for “KEEP DRY”	Symbol for “CAUTION”	Symbol for “CONSULT INSTRUCTIONS FOR USE”
REF Symbol for “CATALOGUE NUMBER”	Symbol for “KEEP OUT OF THE REACH OF CHILDREN”	Symbol for “MANUFACTURER”
Symbol for “DO NOT USE IF PACKAGE IS DAMAGED”	25°C 5°C Symbol for “TEMPERATURE LIMITATION”	Symbol for “KEEP AWAY FROM SUNLIGHT”
MD Symbol for “MEDICAL DEVICE”	IN Symbol for “COUNTRY OF MANUFACTURE”	

Manufactured by: Healix Healthcare Private Limited
110 Manish Industrial Estate No. 4, Navghar Road, Vasai East, Dist. Palghar, Maharashtra 401210, INDIA
Tel: +91-8591106755 | Email: healixhealthcare@gmail.com
LIC. No.: MFG/MD/2023/001052

Kindly retain the instructions for use & packaging box until the complete use of the product / expiry.

The material has been developed solely for professional dental use. Application should be carried out strictly according to the instruction for use. Liability cannot be accepted for damages resulting from failure to observe the instructions of the stipulated area of applications. The user is responsible for testing the material for its suitability and use for any purpose not explicitly stated in this instruction sheet. Description and data constitute no warranty of attributes and are not binding.