PLAQUE DISCLOSING GEL - REVELA - PLAQVIEW GEL

(Dental Plaque Disclosing Gel)

INSTRUCTIONS FOR USE (Please read carefully before use)

Doc. No: **HH/IFU/12** Approved By: **Mr. Vikram Vora** Language: **English**

Revision	01/01
Date	30.08.2026

PRODUCT DESCRIPTION: Plaque Disclosing Gel will differentiate among fresh plaque, mature plaque (more than 48 hours old) and strong acid producing plaque by colour, to aid in patient oral hygiene education. A pink or red colour on the tooth surface indicates fresh plaque accumulation; a blue or purple colour indicates mature plaque that is at least 48 hours old; a light blue colour indicates mature and strong acid producing plaque.

MATERIAL OF CONSTRUCTION: Aqueous gel base, red and blue dyes, benzoate preservatives, thickening agent.

INDICATIONS: Disclosing and differentiation of dental plaque to aid in patient oral hygiene education.

INTENDED USE: For use only by a dental professional in the recommended indications.

PROCEDURE & PRODUCT HANDLING:

- Apply the gel onto the tooth surfaces using a suitable instrument (microbrush, swab or tooth brush). Dispense gel into a dish if you apply on multiple tooth surfaces.
- Following this, instruct the patient to lightly rinse their mouth with water. Identify the plaque by observing the colour of the gel: a pink or red colour indicates fresh plaque accumulation; a blue or purple colour indicates mature plaque at least 48 hours old; a light blue colour indicates mature and strong acid producing plaque.
- Clean the tip of the tube by wiping with a sponge or cotton pellet to remove any excess gel. Close the cap immediately.
- Instruct the patient in proper tooth cleaning and thereby remove any remaining plaque and disclosing gel.

CAUTION / WARNING:

1. Device to be used by a qualified dentist or dental technician only.
2. Avoid contact with eyes. In case of contact, immediately flush with water and seek medical attention.
3. Avoid contact with skin, clothing or equipment. In case of contact, immediately wash with water.
4. Avoid contact with the dental cavity. In case of contact, immediately wash with water; if the area is stained, the colour will be gradually removed.
5. This product contains red and blue dyes, which will stain dental restorative materials.
6. Do not swallow. Do not mix with other products. Keep out of reach of children.
7. If any angioedema symptoms are experienced, this may indicate sensitivity or allergy to the benzoate preservatives. In this event, discontinue the use of the product and refer to a physician.

PRECAUTIONS:

1. Avoid contact with eyes, skin, clothing, equipment and the dental cavity. Do not swallow.
2. Eye contact: immediately flush with water and seek medical attention.
3. Skin contact: immediately wash with water.
4. Clean the tip of the tube after use and close the cap immediately.

CONTRAINDICATIONS:

1. Do not use this material on patients with a sensitivity or allergy to benzoate preservatives.

INTERACTIONS WITH OTHER DENTAL MATERIALS: This product contains red and blue dyes, which will stain dental restorative materials. Do not mix with other products. Kindly communicate and inform the manufacturer with complete details in case of any incidences of interaction not mentioned here.

ADVERSE REACTIONS: Angioedema symptoms may indicate sensitivity or allergy to the benzoate preservatives; in this event, discontinue use of the product and refer to a physician. Kindly communicate and inform the manufacturer in details of any instance.

PACKAGING: Syringe containing 3 g.

STORAGE CONDITIONS: Store in a cool dry low humidity area away from direct sunlight (8–25°C). Keep locked up and out of the reach of children. Close the cap immediately after use. Do not use after expiry date.

SHELF LIFE: 2 years from the date of manufacturing.

DISPOSAL: Dispose in accordance with local, state and federal environmental regulations.

This product is specially formulated for use in dentistry and to be used by qualified dental personnel.

LOT Symbol for “BATCH CODE”	Symbol for “USE BY”	Symbol for “NON STERILE”
Symbol for “KEEP DRY”	Symbol for “CAUTION”	Symbol for “CONSULT INSTRUCTIONS FOR USE”
REF Symbol for “CATALOGUE NUMBER”	Symbol for “KEEP OUT OF THE REACH OF CHILDREN”	Symbol for “MANUFACTURER”
Symbol for “DO NOT USE IF PACKAGE IS DAMAGED”	25°C 8°C Symbol for “TEMPERATURE LIMITATION”	Symbol for “KEEP AWAY FROM SUNLIGHT”
MD Symbol for “MEDICAL DEVICE”	IN Symbol for “COUNTRY OF MANUFACTURE”	

Manufactured by: Healix Healthcare Private Limited
110 Manish Industrial Estate No. 4, Navghar Road, Vasai East, Dist. Palghar, Maharashtra 401210, INDIA
Tel: +91-8591106755 | Email: healixhealthcare@gmail.com
LIC. No.: MFG/MD/2023/001052

Kindly retain the instructions for use & packaging box until the complete use of the product / expiry.

The material has been developed solely for professional dental use. Application should be carried out strictly according to the instruction for use. Liability cannot be accepted for damages resulting from failure to observe the instructions of the stipulated area of applications. The user is responsible for testing the material for its suitability and use for any purpose not explicitly stated in this instruction sheet. Description and data constitute no warranty of attributes and are not binding.